



**MODEL 600040H
HYDRAULIC CADAVER
CARRIER w/ROLLERS**

CAUTION

- Inspect unit and all components for any loosening that may have occurred during shipping •


CAUTION

WHEN TRANSPORTING OR MANEUVERING THE TABLE ALWAYS PLACE THE TABLE IN THE LOWERED POSITION TO PREVENT INJURY.

ATTENTION

Review entire manual before starting assembly





WARNING

Untrained operators can cause injury or be injured. Permit only trained personnel to operate the equipment.

Improper use of the equipment can cause injury. Use the equipment only for the purpose described in this manual.

Improper maintenance can cause injury. Maintain the equipment only as described in this manual.

Improper parts and service can cause injury. Use only prescribed approved parts and service.

Modifying the equipment can cause injury and damage. Use the equipment only as designed by Mortech.

BLOODBORNE DISEASE NOTICE

To reduce the risk of exposure to bloodborne diseases such as HIV-1 and hepatitis when using the equipment, follow the disinfecting and cleaning instructions in this manual.

OPERATING SKILLS AND TRAINING

SKILLS

Operators using the equipment needs

- ☐ A working knowledge of necessary procedures.
- ☐ The ability to carry out necessary service procedures.
- ☐ A complete understanding of the procedures described in this manual.

TRAINING

- ☐ Be trained on the use of the equipment.
- ☐ Read the manual
- ☐ Practice with the equipment before using it in regular service.
- ☐ Be tested on their understanding of the equipment operation.
- ☐ Record their training.

BEFORE USING THE UNIT


Personnel working with this unit needs to read this manual. Assemble of the unit following instructions, and perform any pre-service checks to confirm the units operates properly.


INSPECTING THE UNIT BEFORE USE

- Are all components present?
- Do the moving parts operate smoothly?
- Is unit draining properly?
- Are all nuts, bolts, and pins secured in place?
If unit has an issue contact customer relations.

GENERAL GUIDELINE

Follow standard mortuary-service procedures when working with the unit.



**MODEL 600040H**

Size: 81/205 x 26/66 x 26/74 to
51/129 high without canopy
Design Load: 500 lbs./272 kg.


The Model 600040HR Hydraulic Covered Cadaver Carrier With Rollers allows for the concealed removal and transport of a deceased patient from the hospital room to the morgue. The carrier gives the illusion of being a standard empty gurney while hiding the cadaver beneath a specially fabricated top cover. A roller top frame allows easy transfer of body tray from carrier to lift or carrier to refrigerator. The unit features stainless steel construction and adjustable elevation via foot pedal operated hydraulics.

STANDARD DESIGN FEATURES

- Heavy gauge stainless steel roller frame and cadaver tray
- Hydraulic base with height adjustments via hydraulic foot pedal
- All four 8 in./20 cm. swivel wheels lock instantly with a single foot pedal
- Solid surface base eliminates crevices for liquids to coagulate.
- Sturdy, lightweight aluminum top frame to conceal cadaver and is easily removable
- Custom fabricated cover made of gray launderable cloth
- Hydraulic mechanism elevates cadaver tray to 41.5 in./105 cm.

* All measurements are inches/centimeters





**600040H HYDRAULIC CADAVER
CARRIER w/ROLLER**



**The 600040H in the raised position
(above) In the lowered position.**



The 600040H in the lowered position.



**The 600040H in the raised position
(above) showing the tray with a cloth cover
for concealment.**




**The 600040H in the raised position
(above) with a cloth cover for concealment.**

600040H Hydraulic Cadaver Carrier w/Rollers

Parts listing

**600040H CARRIER
CANOPY ASSEMBLY**





Model 600040H shown with Model M678



TRANSFER FROM LIFT TO REFRIGERATOR

Once the body tray has been transferred to the lifting device, it can then be used to place the body on your storage system or mortuary refrigerator.

Please provide your space available and storage requirements and we will design a Total Roll System to meet your application

We offer a wide range of mortuary refrigerators and storage systems which incorporate the same roller system and allow all equipment to be completely compatible.

If all equipment is compatible, one person can make all body transfers without any manual lifting.

TRANSFER FROM CARRIER TO LIFT

After removal has been completed from the patient room, remove top canopy cover and aluminum frame.

Align the lift with the height of the carrier and simply roll the body tray from the carrier onto the lift roller assembly.

The lift can be used to transfer the body tray to either a storage rack system or a mortuary refrigerator unit.

Model Model M678 Hydraulic Cadaver Lift fitted with our Model M601 Roller Pallet Assembly aligning with a mortuary refrigerator

Hydraulic Cadaver Lift fitted with our Model M601 Roller Pallet Assembly



HYDRAULIC BASE

WARNING

Always use the caster break when a cadaver is on board. Push the carrier to ensure the casters are locked. Injury could result when loading and unloading the carrier.

CAUTION



To avoid damage, remove any equipment that may be in the way before raising or lowering of the carrier. This unit cannot be raised (hydraulics on the base) with a lift under the carrier. The hood may not be used for stepping.


CLEANING

Hand wash all surfaces of the carrier with warm water and mild detergent. Dry Thoroughly. DO NOT STEAM CLEAN, PRESSURE WASH, HOSE OFF OR ULTRASONICALLY CLEAN. Using these methods of cleaning is not recommended and may void this product's warranty. Clean hood storage area regularly. The bottom of the brake rings should be cleaned regularly to prevent wax and/or floor remnants buildup.


PREVENTIVE MAINTENANCE

Preventive Maintenance should be performed at a minimum of annually. A Preventive Maintenance Program should be established for all equipment. Preventive Maintenance may need to be performed more frequently based on the usage level of the product.






OPERATING BASE CONTROLS – SIDE CONTROL



CARRIER OPERATION

Pump pedal (A) to raise the litter.



Depress in the center of pedal (B) to lower both ends of the carrier together.




Pump pedal (A) at the foot end or side of the carrier to raise the litter.

Depress in the center of pedal (C) to lower both ends of the carrier together.


Pedal (E) -Brake and steer function (foot end)

Pedal (F)- Brake and steer function (head end)

RAISING AND LOWERING LITTER HEIGHT-
SIDE CONTROL TO AVOID AN DAMAGE REMOVE ANY
EQUIPMENT THAT MAY BE IN THE WAY BEFORE
LOWERING THE CARRIER.



Pedal (C) - Brake and steer functions (foot end)



Pedal (D)-Brake and steer functions (head end)

WARRANTY

Limited Warranty

Stryker Medical Division, a division of the Stryker Corporation, warrants to the original purchaser for a period of one (1) year after date of delivery. Stryker's obligation under this warranty is expressly limited to supplying replacement parts and labor for, or replacing, at its option, any product that is, in the sole discretion of Stryker, found to be defective. If requested by Stryker, products or parts for which a warranty claim is made shall be returned prepaid to Stryker's factory. Any improper use or any alteration or repair by others in such manner as in Stryker's judgement affects the product material and adversely shall void this warranty. Any repair of Stryker products using parts not provided or authorized by Stryker shall void this warranty. No employee or representative of Stryker is authorized to change this warranty in any way.

Stryker Medical stretcher product are designed for a 10 year expected service life under normal use, conditions, and with appropriate periodic maintenance as described in the maintenance manual for each device. Stryker warrants to the original purchaser that the welds on its stretcher products will be free from structural defects for the expected 10 year life of the stretcher product as long as the original purchaser owns the product.

This statement constitutes Stryker entire warranty with respect to the aforesaid equipment. STRYKER MAKES NO OTHER WARRANTY OR REPRESENTATION EITHER EXPRESSED OR IMPLIED, EXCEPT AS SET FORTH HEREIN. THERE IS NO WARRANTY OF MERCHANTABILITY AND THERE ARE NO WARRANTIES OF FITNESS FOR ANY PARTICULAR PURPOSE. IN NO EVENT SHALL STRYKER BE LIABLE HEREUNDER FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES ARISING FROM OR IN ANY MANNER RELATED TO SALES OR USE OF ANY SUCH EQUIPMENT.

RETURN AUTHORIZATION

Merchandise cannot be returned without approval of Stryker Customer Service Department.

An authorized number will be provided which must be printed on the returned merchandise. Stryker reserves the right to charge shipping and restocking fees on returned items. SPECIAL, MODIFIED, OR DISCONTINUED ITEMS NOT SUBJECT TO RETURN.

DAMAGED MERCHANDISE


ICC Regulation require that claims for damaged merchandise must be made with the carrier within fifteen (15) days of receipt of merchandise. DO NOT EXCEPT DAMAGE SHIPMENT UNLESS SUCH DAMAGE IS NOTED ON THE DELIVERY RECEIPT AT THE TIME OF RECEIPT. Upon prompt notification, Stryker will file a freight claim with the appropriate carrier for damages incurred. Claim will be limited in amount to the actual replacement cost. In the event that this information is not received by Stryker within (15) day period following the delivery of the merchandise, or the damage was not noted on the delivery receipt at the time of receipt, the customer will be responsible for payment of the original invoice in full.

Claims for any short shipment must be made within (30) days of invoice.

INTERNATIONAL WARRANTY CLAUSE

This warranty reflects U.S. domestic policy. Warranty outside the U.S. may vary by country. Please contact your local Stryker Medical representative for additional information.





MAINTENANCE INSTRUCTIONS

MAINTENANCE SCHEDULE

The unit requires regular maintenance and follow the manufactures directions.

DISINFECTING AND CLEANING

Clean all surfaces of the stainless steel with a hard surface disinfectant/cleaner such as Sheila Shine® or SaniZene®. Follow instructions on container.

DISINFECTING AND CLEANING COMPONENTS AND ACCESSORIES

Hand wash with disinfecting soap and warm water. Rinse well with clear water and dry with a towel. (Clean on a regular basis to prolong work life).

WARNING

Improper maintenance can cause injury. Maintain only as prescribed in this manual.

Overall, it is very simple to maintain. If you use for its intended purpose, it will provide many years of reliable service.

CARE OF STAINLESS STEEL

Stainless steel products have a directional #4 brushed finish. This finish is produced using a very fine abrasive cloth. Dragging heavy equipment across the stainless steel surfaces will cause noticeable scratching. Pitting/corrosion can occur when carbon steel products are allowed to remain in contact with the stainless steel in the presence of moisture. (Examples- Steel Wool pads left in the bottom of the sink). Stainless steel can be damaged by exposure to acids.

TYPES OF SURFACE CONTAMINANTS

Dirt - Consist of accumulated dust and a variety of contaminates. Warm water with or without a gentle detergent is sufficient. Next in order are mild non-abrasive powders such as typical household cleaners. These can be used with warm water, bristle brushes, sponges, or cleaning cloths. (Do not use carbon steel brushes or steel wool they may leave particles embedded on the surface which can lead to RUSTING.) For more aggressive cleaning, a small amount of vinegar can be added to the scouring powder. When water contains mineral solids, which leave water spots, it is advisable to wipe the surface completely with dry towels.

Fingerprints and Stains - Fingerprints and mild stains resulting from normal use are the most common surface contaminates. This affect the appearance and seldom have an effect on corrosion resistance. They can be removed with a glass cleaner or by gentle rubbing with a paste of soda ash (sodium carbonate) and water applied with a soft rag. Followed by a thorough warm water rinse and towel dry.

Shop Oil and Grease - These soils may be corrosive and may not allow the surface to maintain passivity, and so removal is a necessity. Soap or detergent and water may be used or a combination of detergent and water plus a solvent.

TYPES OF CLEANERS AND METHODS

General Precautions

Avoid using abrasive cleaners unless absolutely necessary. A "soft abrasive," such as pumice, should be used. Many cleaners contain corrosive ingredients, rinse with clean water.

Clean Water and Wipe - A soft cloth and clean warm water should always be the first choice for mild stains, loose dirt and soils. A final rinse with clean water and a dry wipe will eliminate the possibility of water stains.

Solvent Cleaning - Organic solvents can be used to remove fresh fingerprints, oils and greases that have not had time to oxidize or decompose. The preferred solvent is one that does not contain chlorine, such as acetone, methyl alcohol, and mineral spirits.

EFFECTIVE CLEANING METHODS

Commercial Cleaners - Many commercial cleaners compounded from phosphates, synthetic detergents, and alkalis are available for the cleaning of severely soiled or stained stainless surfaces. When used with a variety of cleaning methods, these cleaners can safely provide effective cleaning.

SCRATCH REPAIR

Surface scratches can be repaired using the following technique. Depending on the severity of the scratch, it may be possible to completely remove it. Sand the scratch using 120 grit emery cloth or paper and firm pressure. Always sand in the direction of the grain. Avoid the natural tendency to sand in an arc, instead sand in a perfectly straight line. Sand until the scratch is gone.

Polish using 3M Scotch Brite pads - Very Fine Grade. Use the same motions as with sanding. Polish until the original finish is restored. Wash and wipe the surface completely with dry towels.

RUST REMOVAL

Rust can be repaired using the following technique. Sand the scratch using 120 grit emery cloth or paper and firm pressure. Always sand in the direction of the grain. Avoid the natural tendency to sand in an arc, instead sand in a perfectly straight line. Sand until the rust is gone.

Polish using 3M Scotch Brite pads - Very Fine Grade. Use the same motions as with sanding. Polish until the original finish is restored. Wash and wipe the surface completely with dry towels.

SEPTEMBER 2010 MFG ISO DOC. #OM600040H (REV A)





600040H Hydraulic Cadaver Carrier w/Rollers

[illegible]

WARRANTY

Mortech Manufacturing warrants all fabrications to be free of defects due to its own workmanship and materials.

Repair and/or replacement of materials furnished that may develop such defects, will be warranted for a period of one year from the date of shipment.

Items not manufactured by Mortech Manufacturing will receive the manufacturer's warranty.

PARTS AND SERVICE

Customer relations and product support are important aspects of Mortech Manufacturing. For assistance with this or any of our fine products please contact us below:

Mortech Manufacturing Company
411 North Aerojet Avenue
Azusa, CA 91702
TEL (626) 334-1471
FAX (626) 334-1704
www.mortechmfg.com
info@mortechmfg.com

© COPYRIGHT MORTECH MANUFACTURING INC., ALL RIGHTS RESERVED.

DISCLAIMER

This manual contains general instructions for the use, operation, and care of this product. The instructions are not all inclusive. Safe and proper use of this product is solely at the discretion of the user. Safety information is included as a service to the user. All other safety measures taken by the user should be within and under consideration of applicable regulations. It is recommended that training on the proper use of this product be provided before using this product in and actual situation.

Retain this manual for future reference. Include it with the product in the event of transfer to new users. Additional free copies are available upon request from Customer Relations.

PROPRIETARY NOTICE

The information in this manual is the property of Mortech Manufacturing. Mortech Manufacturing reserves all patent rights, proprietary design rights, manufacturing rights, reproductions rights, and sales use rights thereto, and to any article disclosed therein except to the extent those rights are expressly granted to others or where not applicable to vendor proprietary parts.

CAUTION

- Inspect unit and all components for any loosening that may have occurred during shipping •

